FossCare™

FOSS

Declaration of Conformity

Manufacturer: FOSS Analytical A/S
Nils Foss Allé 1
DK-3400 Hillerød
Denmark

Instrument:

NIRS™ DS2500 L Metrohm	**60092403**	**91849025**
Name	Part No.	From Serial No.

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Has been produced in compliance with:

ElectroMagnetic Compatibility Directive 2014/30/EU

IEC/EN 61000-6-1:2019	Generic immunity standard. Residential, commercial and light industry
IEC/EN 61000-6-2:2019	Generic immunity standard. Industrial environment
EN 61000-6-3:2012 IEC 61000-6-3:2010	Generic emission standard. Residential, commercial and light industry
IEC/EN 61000-6-4:2019	Generic emission standard. Industrial environment
IEC/EN 61326-1:2013	Electrical equipment for measurement, control and laboratory use Immunity: Basic immunity and Industrial locations Emission: Class B equipment

Low Voltage Directive 2014/35/EU

EN 61010-1:2010 +A1:2016 / IEC 61010-1:2010+A1 2019 Safety requirements for electrical equipment for measurement, control and laboratory use
EN 61010-2-081:2015 / IEC 61010-2-081:2019 Particular requirements for automatic and semi automatic laboratory equipment for analysis and other purposes

RoHS directive 2011/65/EU

And amendments 2015/863/EU on restricted substances (Annex II), amending Directive 2017/2102/EU

REACH Regulation (EC) No. 1907/2006

2020-02-24

Date

Gitte Floridan Strøm
Compliance Manager

ANALYTICS BEYOND MEASURE

FOSS Analytical A/S
Nils Foss Allé 1
DK-3400 Hillerød
Denmark
Phone +45 70 10 33 70
Fax +45 70 10 33 71
E-mail info@foss.dk
Web www.fossanalytics.com

FOSS Analytical Co., Ltd.
6 Louyang Road, Building 1
215121, SIP, Suzhou
P.R. China
Phone +86 512 62 92 01 00
Fax +86 512 62 80 56 30
E-mail info@foss.dk
Web www.fossanalytics.com