

Metrohm AG Ionenstrasse CH-9100 Herisau Switzerland Phone +41 71 353 85 85 Fax +41 71 353 89 01 info@metrohm.com www.metrohm.com

1 EU declaration of conformity

This declaration attests the compliance of the instrument with the standard specifications for electrical instruments and accessories and with the standard specifications for safety and system validation of the manufacturing company.

The responsibility for issuing this EU declaration of conformity lies solely with the manufacturer.

1.1 **Product** validity

This declaration is valid for the following products or product versions:

885 Compact Oven SC

Sample changer with oven for automated processing of small sample series in analytical laboratories.

1.2 Directives

The object of the declaration described above is in conformity with the relevant Union harmonisation legislation:

• 2014/35/EU –Low Voltage Directive, LVD

Directive 2014/35/EU of the European Parliament and of the Council of 26 February 2014 on the harmonisation of the laws of the Member States relating to the making available on the market of electrical equipment designed for use within certain voltage limits; Official Journal of the EU L96, 29 March 2014, p. 357-374

2014/30/EU – EMC Directive, EMC

Directive 2014/30/EU of the European Parliament and of the Council of 26 February 2014 on the harmonisation of the laws of the Member States relating to electromagnetic compatibility; Official Journal of the EU L96, 29 March 2014, p. 79-106

2011/65/EU – Directive for certain hazardous substances, RoHS

Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment; Official Journal of the EU L174, 1 July 2011, p. 88-110

 2012/19/EU – Disposal and recycling of electrical and electronic equipment, WEEE

Directive 2012/19/EU of the European Parliament and of the Council of 4 July 2012 on waste electrical and electronic equipment (WEEE); Official Journal of the EU L197, 24 July 2012, p. 38-71

1.3 Safety specifications

This instrument fulfills the following safety requirements:

Design and type testing	 EN 61010-1: 2010 Safety requirements for electrical equipment for measurement, control, and laboratory use EN 61010-2-010: 2014 Particular requirements for laboratory equipment for the heating of materials EN 61010-2-081: 2015 Particular requirements for automatic and semi-automatic laboratory equipment for analysis and other purposes EN 60529: 2013 – degree of protection IP20 Degrees of protection provided by enclosures (IP Code) ISO 12100: 2010 General principles for design - Risk assessment and risk reduction
	 EN 61140: 2016 – protection class I Protection against electric shock - Common aspects for installation and equipment
Testing in Produc- tion	Every instrument is routine-tested according to EN/IEC 61010-1 Appen- dix F in the production division:

Check of the protective conductor connection and the insulation from power circuits.

1.4 Electromagnetic compatibility (EMC)

Design and type testing

• EN 61326-1: 2013

Electrical equipment for measurement, control and laboratory use - General EMC requirements

1.4.1 Emission

Standards fulfilled

- EN 61000-6-3: 2011
- EN 55011 / CISPR 11: 2016
- EN 61000-3-2: 2014
- EN 61000-3-3: 2013

1.4.2 Immunity

Standards fulfilled

- EN 61000-6-2: 2005
- EN 61000-4-2: 2009
- EN 61000-4-3: 2010
- EN 61000-4-4: 2012
- EN 61000-4-5: 2014
- EN 61000-4-6: 2014
- EN 61000-4-8: 2010
- EN 61000-4-11: 2004
- EN 61000-4-14: 2009
- EN 61000-4-28: 2009

1.5 Manufacturer

Metrohm AG, Ionenstrasse, CH-9100 Herisau/Switzerland

Metrohm AG holds the SQS certificate ISO 9001 – Quality management system for the development, production and sale of instruments and accessories for analyses.

Herisau, Oktober 30th, 2018

P-Hunz

Pr. U. Loha

P. Hunziker

Vice President, Head of Development

Dr. M. Saba Head of Quality Management & Regulatory Affairs

2 Authorizations and certificates

2.1 Authorization Federal Inspectorate for Heavy Current Installations ESTI



This instrument complies with the Swiss Ordinance on Electrical Lowvoltage Equipment (NEV; SR 734.26) and the Swiss Federal Law on Product Safety (PrSG; SR 930.11). The label attests the inspection by the independent national certification institute ESTI, accredited according to ISO/IEC 17065.

The product is listed in the register of authorizations at ESTI.

2.2 IEC certificate



This instrument has been tested in accordance with the IEC standards and certified to the IECEE CB Scheme. The tests also comprise the national differences for Europe, the USA and Canada. Therefore, the instrument is in compliance with the respective standards EN 61010-1, UL 61010-1 and CSA-C22.2 No. 61010-1.

The CB certificate is available in the directory of certified products at electrosuisse.