

**Manufacturer:** FOSS Analytical A/S  
Foss Allé 1  
DK-3400 Hillerød  
Denmark

Instruments:	Name	Part No.	Valid from Serial No.
	<b>XDS Monochromator, Metrohm</b>	<b>60057144</b>	<b>3010-2017</b>
	<b>XDS RapidContent Module, Metrohm</b>	<b>60057052</b>	<b>3013-1487</b>
	<b>XDS RCA with Solids Module, Metrohm</b>	<b>60057051</b>	<b>3013-1487</b>
	<b>XDS MultiVial Module, Metrohm</b>	<b>60057053</b>	<b>3013-1487</b>
	<b>XDS MasterLab Module, Metrohm</b>	<b>60057050</b>	<b>3013-1487</b>
	<b>XDS RapidLiquid Module, Metrohm</b>	<b>60057092</b>	<b>3015-0945</b>
	<b>XDS OptiProbe 1,5m, Metrohm</b>	<b>60057215</b>	<b>3016-0578</b>
	<b>XDS Transmission OptiProbe 1.5m, Metrohm</b>	<b>60057218</b>	<b>3016-0588</b>
	<b>XDS SmartProbe 2m, Metrohm</b>	<b>60057190</b>	<b>3014-0651</b>
	<b>XDS SmartProbe 3m, Metrohm</b>	<b>60057191</b>	<b>3014-0651</b>

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Has been produced in compliance with:

### ElectroMagnetic Compatibility Directive 2014/30/EU

EN 61000-6-1:2007 Generic immunity standard. Residential, commercial and light industry  
 EN 61000-6-3:2007 Generic emission standard. Residential, commercial and light industry  
 EN 61326-1:2013 Electrical equipment for measurement, control and laboratory use.  
 Immunity: Basic requirements  
 Emission: Class B equipment

### Low Voltage Directive 2014/35/EU

EN 61010-1:2010 Safety requirements for electrical equipment for measurement, control and laboratory use.

### Artificial optical radiation Directive 2006/25/EC

### RoHS directive 2011/65/EU

2017-08-08

Date



Per Sonne Sørensen  
Senior Vice President Global Quality

## Dedicated Analytical Solutions

**FOSS Analytical A/S** Tel +45 7010 3370  
Foss Allé 1 Fax +45 7010 3371  
DK-3400 Hillerød E-mail info@foss.dk  
Denmark Web www.foss.dk

**FOSS Analytical Co., Ltd.** Tel +86 512 6292 0100  
6 Louyang Road, Building 1 Fax +86 512 6280 5630  
215121, SIP, Suzhou E-mail info@foss.dk  
P.R. China Web www.foss.dk