

**Manufacturer:** FOSS Analytical A/S  
Nils Foss Allé 1  
DK-3400 Hillerød  
Denmark

**Instrument:** NIRS™ DS2500 Metrohm 60066163 91789106  
Name Part No. From Serial No.

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Has been produced in compliance with:

### ElectroMagnetic Compatibility Directive 2014/30/EU

IEC/EN 61326-1:2013 Electrical equipment for measurement, control and laboratory use.  
Immunity: Basic immunity and Industrial locations  
Emission: Class B equipment

### Low Voltage Directive 2014/35/EU

IEC 61010-1:2010+ A1:2016 / EN 61010-1:2010+A1 2019 Safety requirements for electrical equipment for measurement, control and laboratory use

### Radio Equipment Directive 2014/53/EU

#### Article 3.1a: Health and Safety

IEC / EN 62311:2020 Assessment of electronic and electrical equipment related to human exposure restrictions for electromagnetic fields (0 Hz - 300 GHz)

#### Article 3.1b: EMC

EN 301489-1:2017 EMC standard for radio equipment and services - part 1  
EN 301489-3:2013 EMC standard for radio equipment and services - part 3

#### Article 3.2: Effective use of the radio spectrum

EN 300 330-1:2015 Electromagnetic compatibility and Radio spectrum Matters (ERM); Short Range Devices (SRD)

### Artificial Optical Radiation Directive 2006/25/EC


### RoHS directive 2011/65/EU

And amendments 2015/863/EU on restricted substances (Annex II),  
amending Directive 2017/2102/EU

### REACH Regulation (EC) No. 1907/2006

2020-02-24

Date

  
 Gitte Floridan Strøm  
 Compliance Manager

### ANALYTICS BEYOND MEASURE

**FOSS Analytical A/S** Phone +45 70 10 33 70  
Nils Foss Allé 1 Fax +45 70 10 33 71  
DK-3400 Hillerød E-mail info@foss.dk  
Denmark Web www.fossanalytics.com

**FOSS Analytical Co., Ltd.** Phone +86 512 62 92 01 00  
6 Louyang Road, Building 1 Fax +86 512 62 80 56 30  
215121, SIP, Suzhou E-mail info@foss.dk  
P.R. China Web www.fossanalytics.com