



dSolvotrode

6.00203.300

Digital, combined pH electrode for OMNIS for all non-aqueous acid/base titrations. The glass membrane is optimized for poorly conducting solutions and thanks to the flexible ground-joint diaphragm, the electrode is suitable for contaminated samples.

This electrode can be used with non-aqueous reference electrolytes (lithium chloride or tetraethylammonium bromide).

Storage in corresponding reference electrolyte.
dTrodes can be used on OMNIS Titrators.

Below, the accessories are grouped into Scope of delivery and Optional accessories.
Please keep this printout at hand for ordering replacement material.
These lists may be subject to change.

Optional accessories

Order no.	Description
-----------	-------------

6.02104.300 Electrode cable plug-in head Q / plug P, 0.55 m

Cable for connecting a dTrode (plug-in head Q) to the digital measuring module from OMNIS.



6.02104.310 Electrode cable plug-in head Q / plug P, 1.5 m

Cable for connecting a dTrode (plug-in head Q) to the digital measuring module from OMNIS.



6.2320.000 TEABr 0.4 mol/L in ethylene glycol (250 mL)

Electrolyte solution TEABr (tetraethylammonium bromide in ethylene glycol), $c(\text{TEABr}) = 0.4 \text{ mol/L}$

