



## Ag/AgCl reference electrode (length 12.5 cm)

6.0726.100

Silver/silver chloride reference electrode with Double-Junction System, installation length 10 cm. The standard ground-joint 14/15 enables easy assembly, and the flexible ground-joint diaphragm, which is insensitive to contamination, can be replaced at any time. The reference electrolyte and the bridge electrolyte can be selected freely according to use, and are easy to replace. This sensor is supplied without electrolyte filling.

Below, the accessories are grouped into Scope of delivery and Optional accessories. Please keep this printout at hand for ordering replacement material. These lists may be subject to change.





### Scope of delivery 6.0726.100

Qt.	Order no.	Description
-----	-----------	-------------

Together with 6.2008.050 storage vessel holder, it provides a support for the electrode on 807 Dosing Units.



## Optional accessories

Order no.	Description	
6.1243.030	Diaphragm for reference electrode  Spare ground-joint diaphragm for 6.0726.1XX and 6.0729.1XX reference electrodes	
6.2106.020	Strand / 1 m / 2 x B  Electrode cable with 2 x plug B (4 mm) for connecting electrodes with plug-in head B	
6.2106.060	Strand / 2 m / 2 x B  Electrode cable with 2 x plug B (4 mm) for connecting electrodes with plug-in head B	
6.2308.020	Electrolyte 3 mol/L KCl (250 mL)  Electrolyte solution $c(\text{KCl}) = 3 \text{ mol/L}$ , (for Ag/AgCl reference systems)	

6.2310.010 Electrolyte KNO<sub>3</sub>-1 mol/L 250 mL

Electrolyte solution KNO<sub>3</sub> 1 M (reference electrolyte for combined silver electrodes)



---

6.2320.000 TEABr 0.4 mol/L in ethylene glycol (250 mL)

Electrolyte solution TEABr (tetraethylammonium bromide in ethylene glycol), c(TEABr) = 0.4 mol/L

